

Metal Implants

This invention relates to metal implants for use in surgical procedures, and in particular to the
5 introduction of a biocidal material into such implants to suppress or control infection.

Various surgical procedures require the use of implants. For example cancerous bone may be removed, in
10 prosthetic surgery, to be replaced by a metal implant. Such an implant may for example be of titanium alloy, which is very strong and relatively light. To ensure a hard-wearing surface the provision of a titanium nitride coating has been suggested. There is furthermore a risk
15 of introducing infection when implanting such metal implants, and it has been suggested that metallic silver might be electroplated onto metal implants, the silver being a biocidal material that can control infection without causing toxic effects to the patient. However
20 such coatings, whether of titanium nitride or silver, may be undercut due to corrosion from body fluids, so that the coating may detach from the implant, which may can increase wear and cause tissue damage.

25 According to the present invention there is provided an implant for use in a surgical procedure, the implant comprising a metal substrate and a surface layer that is integral with the metal substrate, the layer incorporating a biocidal metal deposited from a solution.
30

The invention also provides a method of producing such an implant.

Such an integral surface layer may be generated by
35 growing the layer from the metal itself, for example by an anodising process; or alternatively by depositing the

layer for example by electroplating, followed by diffusion bonding so that the layer becomes integral with the metal of the implant. Anodising forms an adherent oxide layer, although if it is carried out in phosphoric acid then a phosphate may be formed. Such an adherent phosphate layer may also be modified to form a hydroxyapatite layer, which can stimulate bone growth.

The biocidal material should preferably be effective for at least 6 weeks, preferably for up to 6 months after surgery, and the release rate should be low to avoid toxic effects on body cells. Furthermore the total quantity of biocidal material is preferably also limited to minimize any toxic effects.

It is also desirable if the surface is highly polished before production of the surface layer. This may for example be achieved by electropolishing.

In principle, a range of different metals may be used for the biocidal metal. In particular, if the layer is a metal layer deposited by electroplating then it clearly must be stable to corrosion. Gold, platinum, iridium and palladium would be potentially suitable, although expensive; silver is preferable as it is not particularly soluble in body fluids due to the presence of chloride ions and the low solubility of silver chloride. If the surface layer contains the biocidal metal in ionic form, then a wider range of metals would be possible. In addition to the elements already mentioned, copper, tin, antimony, lead, bismuth and zinc might be used as ions combined into an insoluble matrix for example of metal oxide or metal phosphate. The rate of release would be controlled, in this case, primarily by the strength of the absorption of the metal ions in the matrix.

- 3 -

The metals that may be used to make such prosthetic implants are typically a form of stainless steel, a titanium alloy, or a cobalt/chromium alloy, although zirconium could also be used. The standard alloys for this purpose are titanium 90% with 6% aluminium and 4% vanadium (British standard 7252), or chromium 26.5-30%, molybdenum 4.5-7%, and the remainder cobalt (British standard 7252 part 4).

10 Preferably the implant is initially polished to provide a very smooth surface. Both stainless steel (chromium/iron/nickel) and cobalt/chromium alloy can be electro-polished using as electrolyte a mixture of phosphoric acid and glycerine, or a mixture of phosphoric
15 acid and sulphuric acid. Titanium alloy can be electro-polished using acetic acid, or a mixture of nitric and hydrofluoric acids. Alternatively the implants might be subjected to a combination of anodic passivation with mechanical polishing, which may be referred to as
20 electrolinishing, this process removing the oxide that protects surface roughness, the surface at that point then being electrochemically re-passivated, so producing a mirror-smooth finish. Various electrolytes are suitable for this purpose, including nitric acid mixed
25 with sulphuric acid, sodium hydroxide, sodium phosphate, or sodium hydroxide mixed with sodium nitrate.

After polishing the surface of the metal, either silver deposition or surface conversion can take place.
30 Considering surface conversion first, a layer of metal oxide or phosphate may be formed by anodising in a suitable electrolyte, so that the oxide or phosphate layer builds out from the surface of the metal. Biocidal metal ions can then be absorbed from an aqueous salt
35 solution into the oxide or phosphate matrix, for example the ions Ag^+ or Cu^{++} . Cations of palladium, platinum or

even ruthenium could be absorbed in a similar way. If desired, deposited silver, platinum or palladium ions could then be converted to metal, or deposited ruthenium ions converted to insoluble RuO_2 , within the oxide or
5 phosphate surface coating, this reaction being performed chemically or electrochemically or by light.

Considering now silver deposition, the coating should be thin to prevent toxic effects. A high degree
10 of adherence to the underlying metal can be ensured by first removing the surface oxide layer by anodic etching, followed by a brief reversal of polarity in the presence of appropriate ions, so as to cover the surface with a thin coating of silver. This may be repeated to ensure
15 there are no pin-holes. The plating electrolyte may include hydrofluoric acid, or may be an alkaline cyanide electroplating electrolyte. After deposition, the silver coating should be diffusion bonded so as to form an inter-metallic layer, by heating the implant to an
20 elevated temperature. Typically it should be heated to above 800°C , preferably between 810°C and 950°C , in an inert atmosphere for example of argon for a period of between 1 and 6 hours. This substantially eliminates the risk of coating delamination. However with titanium-
25 based implants the temperature must not exceed 850°C as titanium would undergo a phase change from alpha to beta form above this temperature.

In place of silver, other metals such as platinum or
30 palladium may be electro-deposited and then thermally treated in a similar fashion so as to form an inter-metallic layer.

The invention will now be further and more
35 particularly described, by way of example only.

- 5 -

A hip implant is made of titanium alloy (Ti/Al/V). The implant is cleaned ultrasonically using first acetone as the liquid phase, and then a 1 M aqueous solution of sodium hydroxide, and is then rinsed in de-ionised water.

5 The cleaned implant is then immersed in a stirred 12 weight % solution of phosphoric acid, and is anodised for 2 hours at a maximum voltage of 10 V and a maximum current of 10 mA/cm², so as to form a surface coating of titanium phosphate. It is then rinsed in de-ionised

10 water again. The surface, which is initially pale grey, turns to a darker matt grey as a consequence of the anodising, with a slightly yellow hue.

The implant is then immersed in a stirred 0.1 M

15 aqueous solution of silver nitrate, and left for 2 hours. As a result of ion exchange there is consequently some silver phosphate in the titanium phosphate coating. The implant is then ready to be implanted. During exposure to body fluids there will be a slow leaching of silver

20 ions from the phosphate layer, so that any bacteria in the immediate vicinity of the implant are killed. Infection arising from the implant is therefore suppressed.

25 Experimental samples of this titanium alloy were cleaned, anodised to form a layer of titanium phosphate, and then subjected to ion exchange to form silver phosphate, following the procedure described above. One sample was placed in direct daylight for 110 hours; the

30 exposed surface became darkened as a result of this exposure to daylight, indicating the formation of silver metal by photo-reduction. The other sample was immersed in a solvent containing a mixture of 4 M nitric acid and 0.5 M sodium fluoride (equivalent to hydrofluoric acid)

35 to dissolve the coating. The dark grey surface coating was removed completely within 3 minutes, leaving a

silver-grey finish. The resulting solution was analyzed for the presence of silver by atomic absorption spectrometry, and the concentration of silver was found to be equivalent to an average surface loading of 735 $\mu\text{g}/\text{cm}^2$.